

Calibration Certificate



Manufacturer: PACIFIC CA SYSTEMS
Unit Model: O2 / CO2 Analyser
Customer:
Carmel Electrics

Service Number: 331572357
Serial Number: PDA2031
Unit Number: 1
Job Number: GS 42472
Date Service d: 16-Jul-10
Customer Ref:

info@gastech.com 24 Baretta Rd
Tel:+61 8 6108 0000 WANGARA WA 6065
Fax:+61 8 9408 1868 ABN 64 087 801 634

Warranty Claim Procedure:

Return unit freight pre-paid for warranty repairs.

If Unit Number is removed, warranty claims must be accompanied by last Calibration Certificate.

Next Service: 16-Jul-11

Unit Report

Condition received: Fair

Comments on receiving unit

Due for service and calibration, O2 sensor date expired, pump faulty.

Comments on completing unit

Parts replaced, unit serviced and calibrated.

Functional Tests

Alarm operation: Not Applicable **Display operation:** Checked
CircuitBoard: Checked **Battery test:** Checked

Sensor Report

Test/Calibration: PASSED

| | | | | |
|---|------------------------------|---|---|---|
| Date Cell Fitted 16/07/2010 | Cell Number 9608 | Zero: Set at 0 % Span: Set at 2.0 % Alarm: Not Applicable | Certificate Number NATA#88866/1 | Span Gas Concentration 1.97% |
| SO2 Date Cell Fitted 16/07/2010 | Cell Number 100601 | Zero: Set at 2.0 % Span: Set at 20.9 % Alarm: Not Applicable | Certificate Number Lot# 974446 Cyl BC16777 | Span Gas Concentration 10 ppm |
| SO2 Date Cell Fitted 10/02/2014 | Cell Number 1234 | Zero: Set at 0 ppm Span: Set at 27 % Alarm: | Certificate Number Lot# 974446 Cyl BC16777 | Span Gas Concentration 10 ppm |

Page 1 of 1 (End of Report)

Serviced by: Steve Bond

Signature

